



ALARM

Titel ABCDEFG: THi-1

Additional Information

Line 1:
 Line 2:
 Line 3:
 Line 4:
 Line 5:
 Line 6:
 Line 7:
 Line 8:

Device Configuration

Type:	Libero THi1 V3.16	Inspection Range:	All data
Logger ID:	18000287	Current State:	Logging Transit
Log Interval / Duration:	5 m / 27.8 d	Remaining Battery:	3.0 y
Log Mode:	Start/Stop	Logger Start:	Feb/08/2011 09:17:21
Report Time Base:	GMT +02:00	Checksum:	Ab1C / 270.132.801
Configured by:	S191, EC66/ELPRO, Feb/07/2011 21:49:54		

Alarm Conditions	Total Time	Status
T: Upper Threshold: 30.0 °C	T: Time above Threshold: 20 m	ALARM
T: Lower Threshold: 15.0 °C	T: Time below Threshold: 7.8 h	ALARM
rH: Upper Threshold: 60.0 %rH	rH: Time above Threshold: 13.9 h	ALARM
rH: Lower Threshold: 30.0 %rH	rH: Time below Threshold: 7.8 h	ALARM
Alarm Delay: 10 m		

Logging Results

Highest Temperature:	45.2 °C; Feb/08/2011 22:37:21	Transit Start at:	Feb/08/2011 09:17:21
Lowest Temperature:	-25.3 °C; Feb/08/2011 22:47:21	Arrived at:	not available
Average Temperature:	16.2 °C	Alarm at:	Feb/08/2011 09:27:21
MKT	20.5 °C	File created:	Feb/09/2011 10:43:17
Highest rel. Humidity:	100.0 %rH; Feb/08/2011 22:37:21		
Lowest rel. Humidity:	7.0 %rH; Feb/08/2011 22:47:21		
Average rel. Humidity:	51.2 %rH		

